



Analytical Industries Inc.
Advanced Instruments Inc.

PSR-11-39-JJ



Replacement for JJ Rebreather. More...

Advanced galvanic type % oxygen sensor with excellent stability and accuracy under stringent applications. All rebreather % oxygen sensor are subjected to the most extensive stability test, output in air, output at 100% O₂ and output reading at 100% O₂ at 1.8 ATA.

The widest range of oxygen sensors offered by Analytical Industries, Inc. are "Made in USA"

OEM Equipment:

JJ Rebreather "Retrofit"
 R-17JJ-CCR Sensor

TECHNICAL SPECIFICATIONS

Measuring Range	0-100%
Accuracy ¹	±2% of Full Scale
Signal Output ²	11 - 13 mV
Linearity	±2% of Full Scale
Response T90	6 sec
Temp Coefficient	compensated
Operating Temp	0 to 45°C
Recommended Storage ³	0 to 25°C
Qualified at	1.8 ATA
Shelf Life ⁴	6 months
Humidity Non-condensing	0-99% RH
Expected Life	42 months
Warranty ⁵	12 months
Electrical Conn	90° SMB Female on 5.25" cable

Conditions: Specification validated during design and in pursuit of improvement are subject to change without notice 1) At constant temperature and pressure. 2) In air (20.9% oxygen) at 25°C and 1 atm. 3) Sensor may be stored up to 55°C on an intermittent basis, for example, during transportation. 4) In original Package at 25°C and 1 atm. 5) Under normal operating conditions, the sensor is warranted to be free of defects in material and workmanship for the specified period provided the sensor is properly installed and operated. The sole remedy for sensor determined to be defective by Analytical Industries Inc. is limited to replacing the sensor. Analytical Industries Inc. will not be liable for buyer's negligence, misapplication, abuse or accident.